

SUPPLEMENTARY DATA

Supplementary Figure 1. Bland-Altman plot of Pulse Wave Velocity actual and quality control (QC) data on 54 participants.

Actual PWV: mean (SD) = 5.68 (1.04) m/sec; Duplicate PWV: mean (SD) = 5.73 (1.13) m/sec; Average difference: mean (SD) = -0.0435 (0.526), $p = 0.5459$.

